

Acute Changes in Mitral Valve Geometry after Percutaneous Valve Repair with MitraClip XTR by Three-Dimensional Echocardiography

Maurizio Tusa¹, Marta Barletta¹, Antonio Popolo Rubbio¹, Nicole Travaglio¹, Silvia Saffioti¹, Giuseppina Granata¹, Valentina Mantovani¹, Elisa Stefanini¹, Anca Corciu¹, Luca Testa¹, and Francesco Bedogni¹

¹IRCCS Policlinico San Donato

March 19, 2021

Abstract

BACKGROUND. Mitral valve (MV) repair with MitraClip system is a safe treatment option for high-risk patients with significant mitral regurgitation (MR). We aimed to characterize, by three-dimensional echocardiography (3D-E), changes occurring in MV after implantation of third generation MitraClip XTR device, with specific reference to the underlying MR mechanism (functional vs. degenerative, FMR vs DMR). **METHODS.** We prospectively enrolled 59 patients, who underwent intra-procedural 3D-E before and after device deployment. Three-D datasets were analysed off-line, using a dedicated semiautomatic software, to obtain parametric quantification of mitral anatomy. **RESULTS.** Post-procedural MR of mild or lesser degree was achieved in 40 patients (68%), with no differences between FMR and DMR (p 0.9). After MitraClip XTR implantation, the FMR group experienced an immediate annular resizing, with reduction of antero-posterior diameter (p 0.024) and sphericity index (p 0.017), next to a recovery of physiological saddle-shape, defined by lower non-planar angle (p [?] \leq 0.001) and higher annulus height to commissural width ratio (p [?] \leq 0.001). On the opposite, the DMR group revealed a significant decrease of maximum annular velocity (p 0.027), addressing a mechanic effect of the device deployment. Finally, baseline anterior mitral leaflet angle was found as the strongest independent predictor of acute procedural result (OR 0.91, [CI 0.84-0.99], p = 0.030). **CONCLUSIONS.** MitraClip XTR implantation acts in restoring the original mitral geometry, with distinctive effects according to MR mechanism. Three-D parametric quantification of MV sheds new light on changes occurring in the valvular apparatus, and helps identifying possible new predictors of acute procedural success.

Acute Changes in Mitral Valve Geometry after Percutaneous Valve Repair with MitraClip XT_R by Three-Dimensional Echocardiography

3D Mitral Geometry after MitraClip XT_R deployment

Maurizio B. Tusa¹, MD; Marta Barletta¹, MD; Antonio Popolo Rubbio¹, MD; Nicole Travaglio², MD; Silvia Saffioti³, MD; Giuseppina Granata¹, MD; Valentina Mantovani¹, MD; Elisa Stefanini¹, MD; Anca I. Corciu¹, MD; Luca Testa¹, MD, PhD; Francesco Bedogni¹, MD

1. IRCCS Policlinico San Donato, Department of Cardiology, San Donato Milanese - Milan, Italy

E-mail: tusa.maurizio@libero.it, marta.barletta@outlook.it, antoniorubbio@yahoo.it,
nicole.travaglio@hotmail.it, silviasaffioti@gmail.com, giusy.granata@gmail.com,
valentina.mantovani84@gmail.com, elisa.stefanini@hotmail.it, ancabistrita@yahoo.com, luctes@gmail.com,
francesco.bedogni@grupposandonato.it

Address for correspondence:

Maurizio Barnaba Tusa

Address: Piazza Edmondo Malan, 2, 20097 San Donato Milanese MI

E-mail: tusa.maurizio@libero.it

Telephone: +39 3389307149

FAX: 02 52774757

Hosted file

2. ABSTRACT (Echocardiography).pdf available at <https://authorea.com/users/402665/articles/514457-acute-changes-in-mitral-valve-geometry-after-percutaneous-valve-repair-with-mitraclip-xtr-by-three-dimensional-echocardiography>

Hosted file

3. MANUSCRIPT (Echocardiography).pdf available at <https://authorea.com/users/402665/articles/514457-acute-changes-in-mitral-valve-geometry-after-percutaneous-valve-repair-with-mitraclip-xtr-by-three-dimensional-echocardiography>

Hosted file

4. TABLES & FIGURES (Echocardiography).pdf available at <https://authorea.com/users/402665/articles/514457-acute-changes-in-mitral-valve-geometry-after-percutaneous-valve-repair-with-mitraclip-xtr-by-three-dimensional-echocardiography>




Figure 2




Figure 3

