

Table 3: Linear calibrating equations.

TB1	TB2	TB3	TB4
$V=12.02\ N$	$V=68.11\ N$	$V=155.98\ N$	$V=774.50\ N$

Where: V is the volume measured by the TB and N is the tipping number register by the datalogger over a time interval.